3M™ Attest™ 290/290G Auto-readers

Accurate, Reliable Easy-to-use

Easy to operate

No calibration: The 3M Attest 290/290G Auto-readers do not require time-consuming calibration.

Automatically reads indicator results: A red (+) light indicates a sterilization failure. A green (–) light means that the sterilization conditions have been met.

Alarms: For improved monitoring, an audible alarm sounds and the positive red (+) light illuminates when a positive result is detected. An alarm also sounds if a 3M™ Attest™ Rapid Readout Biological Indicator is removed from the reader before final results have been detected.

Maintenance-free

The 3M Attest 290/290G Auto-readers do not require any routine maintenance.

3M™ Attest™ Rapid Readout Biological Indicator System

For use with 3M Attest Rapid Readout Biological Indicators and Test Packs

3M Attest 290 Auto-reader for Steam Sterilization

- 3M™ Attest™ Rapid Readout Steam Biological Indicators 1291 and 1292
- 3M™ Attest™ Rapid Readout Biological Indicators for Steam Test Packs 1296 and 41382

3M Attest 290G Auto-reader for Ethylene Oxide Sterilization

- 3M™ Attest™ Rapid Readout Biological Indicator for Ethylene Oxide 1294
- 3M™ Attest™ Rapid Readout Test Pack for Ethylene Oxide 1298

3M™ Attest™ 290 Auto-reader for Steam

Accurate, reliable and easy-to-use units for reading 3M™ Attest™ Rapid Readout Biological Indicators for steam sterilization.

Catalog Number	Description	Packaging
290	3M Attest Auto-reader for Steam *For use with 3M Attest Rapid Readout Biological Indicators and Test Packs for Steam 1291, 1292, 1296, 1296F, 41382 and 41382F*	1 each/case

Associated Products, for use with the 3M Attest 290 Auto-reader for Steam

1291 and 1292
3M™ Attest ™ Rapid Readout Steam Biological Indicators

1296 and 1296F
3M™ Attest™ Rapid Readout Steam Test Packs

41382 and 41382F
3M™ Attest™ Rapid 5 Steam-Plus Test Packs

3M™ Attest™ 290G Auto-reader for Ethylene Oxide

Accurate, reliable and easy-to-use units for reading 3M™ Attest™ Rapid Readout Biological Indicators for Ethylene Oxide.

Catalog Number	Description	Packaging
290G	3M Attest Auto-reader for Ethylene Oxide *For use with 3M Attest Rapid Readout Biological Indicators and Test Packs for Ethylene Oxide 1294, 1298 and 1298F*	1 each/case

Associated Products, for use with the 3M Attest 290G Auto-reader for Ethylene Oxide

1294
3M™ Attest ™ Rapid Readout Biological Indicator for Ethylene Oxide

1298 and 1298F
3M™ Attest ™ Rapid Readout Test Pack for Ethylene Oxide

Warranty

One year on the Attest Auto-readers, two years on the bulb.

3M™ Attest™ Monitoring Interface for SPM® Software

Automatic, real-time documentation of 3M Attest biological indicator (BI) results.
With SPM® Software and the 3M™ Attest™ Monitoring Interface, you don't have to worry whether your busy staff has documented the results of a BI test. The results are delivered automatically, and, if positive, alerts are communicated in real-time.

To learn more about the 3M™ Attest™ Auto-readers, call the 3M Health Care Helpline at 1-800-228-3957 or visit www.3M.com/infectionprevention.

3M

3M Health Care
Infection Prevention Division
3M Center, Building 275-4W-02
St. Paul, MN 55144-1000
USA
1-800-228-3957
www.3M.com/infectionprevention

Please recycle. Printed in U.S.A.
SPM is a registered trademark of Materials Management Microsystems.
3M and Attest are trademarks of 3M Company.
© 3M 2009. All rights reserved.
70-2009-5995-8 (89.5) DPI